



SERVICE BULLETIN

KODAK X-OMAT Processor

Eastman Kodak Company, Health Imaging, Rochester, New York 14650

SERVICE BULLETIN NO. 260

March 1997

3434 KODAK X-OMAT 3000 RA Processor

Announcing Design Improvements to the Developer Rack and Updated Parts List

A number of changes have been made to the DEVELOPER RACK in the *Kodak X-Omat* 3000 RA to improve film quality. The changes include new GEARS, DRIVE CHAIN and SPROCKETS that provide a more even transport motion of the film to reduce artifacts. The changes have undergone significant testing to assure their performance and life.

The changes have been phased in over a period of time. As a result of the phasing-in period there are currently 4 different DEVELOPER RACKS in the field.

A new Parts List has been developed which includes all 4 phases of the DEVELOPER RACK. It also contains a section to help you identify which phase DEVELOPER RACK you have. The other sections of the Parts List have also been updated. The Parts List, Part Number 5B6336, dated February 1997 is available from Service Parts Management and Advertising Distribution.

Dave Green
Service Engineer
Customer Equipment Services

For More Information:

U.S. Distributors Contact:
Health Imaging
Customer Support Operations
1-800-336-4722

CES Personnel Contact:
Film Handling TAC
U.S.: 1-800-822-1414
Canada: 1-800-433-1414
International: 716-724-4675

Kodak and X-Omat are trademarks.

9B8933

