



SERVICE BULLETIN

Kodak M35 A-M and M35 M X-Omat Processors

Eastman Kodak Company, Health Imaging, Rochester, New York 14650

SERVICE BULLETIN NO. 265

JULY 1997

Kodak M35 A-M X-Omat Processor, Service Code 3133 ***Kodak M35 M X-Omat Processor, Service Code 3408***

Changes to the DEVELOPER RACK

The DEVELOPER RACK for following Processors will be manufactured with smooth GUIDE SHOES, 7C7860 (Replaces GUIDE SHOES 1C7635).

The new DEVELOPER RACK is 7C7862 (Replaces DEVELOPER RACK, 1C7315).

Processor M35 A-M, Serial number 53245 and above.

Processor M35 M, Serial number 76996 and above.

For More Information:

U.S. Distributors Contact:

Health Imaging
Customer Support Operations
1-800-336-4722

GCSS Personnel Contact:

Film Handling TAC
U.S.: 1-800-822-1414
Canada: 1-800-433-1414
International: 716-724-4675

Publication Change Notice Table

Print Date	Pub. No.	ECO No.	Affected Pages	File Name	Notes
July 1997	7C7881	2604-116	All	sb3133_jul97_265.doc	new publication

Kodak and X-Omat are trademarks.

7C7881

sb3133_jul97_265.doc

Printed in USA

EASTMAN KODAK COMPANY ROCHESTER, N.Y. 14650

HEALTH IMAGING

