



SERVICE BULLETIN

KODAK X-OMAT Processor

Eastman Kodak Company, Health Imaging Division, Imaging Service Center, Rochester, New York 14650

SERVICE BULLETIN NO. 215

March 1995

KODAK X-OMAT Multiloader 300 KODAK X-OMAT 270 RA Processor KODAK X-OMAT 180 LP Processor

New Springs Available for the Dryer Rack, Multiloader 300 and 270 RA Processor Only

Replacement of these Springs in the Dryer Racks is optional, based on the service history of the Processor. The new Springs are made of a special, high-quality stainless steel alloy, which will greatly reduce corrosion of the Springs. See the illustrations of the Dryer Rack in the Parts List to locate the Springs. Order from Service Parts Management.

Processor	Old Part No.	New Part No.
Multiloader 300 and 270 RA	246580	5B6165

New Bulkhead Fittings Available



The Bulkhead Fittings for these 3 Processors now have Reinforcements. The new design will make the Fittings stronger, which will reduce the chance of solution leakage due to broken or cracked Fittings. See the illustrations on the next page.

Fittings that were installed in Modifications No. 3 (on the 180 LP Processor), No. 16 (on the 270 RA Processor), and No. 23 (on the Multiloader 300) **are acceptable**. It is not necessary to install the new Fittings on these Processors unless there is leakage.

Order new Fittings from Service Parts Management. See the table below and the illustrations on the back of this bulletin.



Processor	Description of Fitting	New Part No.
Multiloader 300	3/8-in. Elbow	1C7115
	5/8-in. Elbow	1C7114
270 RA	3/8-in. Straight	1C7113
180 LP	5/8-in. Straight	1C7112

Existing Fittings



H118_0053CA

New Fittings



H118_0054CA

For more information, please contact:

Eastman Kodak Company

Health Imaging Division

Imaging Service Center

343 State Street, 3rd fl, Bldg 20

Rochester, NY 14650-1131

(800) 336-4722

Please be prepared to give your Customer Number and your equipment serial number.

Kodak and X-Omat are trademarks.
N-928

